

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/2000-NL1-03.03

Project number 211292

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: P.P.M. van Enckevort

Applicant

Name: ASCELL SENSOR, S.L.
Address: Jaune Llorens i Vidal, 46
08757 Corbera de Llobregat
Barcelona
Spain

Manufacturer of the certified pattern

Name: ASCELL SENSOR, S.L.
Address: Jaune Llorens i Vidal, 46
08757 Corbera de Llobregat
Barcelona
Spain

Identification of the certified pattern

Type : BM....
Fraction : $P_i = 0.7$
Temperature limit : 40 °C / -10 °C

Maximum capacity (E_{max})	10 kg up to and including 1800 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	10 000


Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/2000-NL1-03.03
Project number 211292
Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC6178 and the appertaining documentation folder), with the requirements of the following Recommendation(s) of the International Organisation of Legal Metrology (OIML):

R60
edition 2000 (E)
for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report:
N° R60/2000-NL1-03.03A, that includes 40 pages
N° R60/2000-NL1-03.03B, that includes 38 pages
N° R60/2000-NL1-03.03C, that includes 37 pages.

The issuing authority
P.P.M. van Enckevort
Manager Certification Delft

21 January 2003

The OIML member
G.J. Faber

21 January 2003

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.